


## Sentinel<sup>™</sup> Tunnel Drill Bit

## To Use:

- Access auxillary anteromedial portal with the knee flexed
- Manually advance drill through the skin incision (AM Portal) with the sharp end (cutting flute) facing away from the medial femoral condyle
- Vertical black line on base of the drill should face surgeon and away from condyle
- Hold drill so it does not spin
- Push past PCL
- Place drill bit head directly on the articular surface of the lateral femoral position







This unique drill bit features an eccentric mono-fluted cutting edge for drilling bone tunnels in ACL and PCL reconstruction procedures.

Cannulated for use with a 2.4mm High Strength Guide Pin

Rounded side designed to protect cartilage and soft tissue prior to drilling Eccentric monofluted cutting edge

S8513

9744

C8675



Description	Catalog Number
Sentinel, 9" total working length, sterile, single use drill bit (4 per box)	
5.5mm	\$8555
6.0mm	\$8560
6.5mm	\$8565
7.0mm	\$8570
7.5mm	\$8575
8.0mm	\$8580
8.5mm	\$8585
9.0mm	\$8590
9.5mm	\$8595
10mm	\$8510
10.5mm	\$85105
11mm	\$8511
11.5mm	\$85115
12mm	\$8512
12.5mm	\$85125

CONMED CORPORATION PRODUCT AREAS:

ARTHROSCOPY • ELECTROSURGERY • ENDOSCOPY • ENDOSURGERY • GASTROENTEROLOGY • INTEGRATED SYSTEMS • PATIENT CARE • POWERED INSTRUMENTS • PULMONOLOGY

13<sub>mm</sub>

14" High Strength Guide Pin, 2.4mm dia.

Graft Passing Guide Pin, 2.4mm dia.



11311 Concept Boulevard Largo, FL 33773-4908 (727) 392-6464

Customer Service: 1-800-237-0169

FAX: (727) 399-5256

International FAX: +1 (727) 397-4540 email: customer service@linvatec.com

www.linvatec.com

© 2008, Linvatec Corporation, a subsidiary of CONMED Corporation, 10/2008, CST 1029