

Getinge Assured

Emulating Indicator 4

The Getinge Assured Emulating Indicator 4 is designed for use in steam sterilizers operating at either 121°C for 12 minutes or 134°C for 4 minutes.

The Emulating Indicator gives visible indication that sterilizing parameters (time, temperature and steam) were met according to EN ISO 11140-1.



Performance Characteristics

The Getinge Assured Emulating Indicator 4 diamond symbol changes color from purple to green when exposed to appropriate steam sterilization conditions.

Stated Values

Time at 134°C	Time at 121°C
4 minutes	10 minutes

*As determined in a steam sterilization resistometer. Resistometer end point times may be different than standard steam sterilizer endpoint times.

Physical Characteristics

- **Lamination:** Top and bottom
- **Dimensions:** L 101mm x W 19mm
- **Packaging:** 250 indicators in a zip lock moisture barrier bag

Environmental Limitations

Storage: In provided sealed packaging at room temperature 10°-38°C and 10-60% Relative Humidity. Keep away from light and heat exposure.

Shelf Life

24 months. The expiry date is printed on the product packaging.

Standards Classification

EN ISO 11140-1, Type 6.

Ordering Information

Getinge Assured
Emulating Indicator 4

Order Number
504051000