

Getinge Assured

Helix Test Formaldehyde

The Getinge Assured Helix Test Formaldehyde verifies air removal, formaldehyde exposure and deep vacuum achievement necessary for the sterilization of hollow lumen instruments.

The Helix Test Formaldehyde is suitable for daily sterilizer monitoring. The test is designed to be used in each sterilization cycle as an independent control device.



Performance Characteristics

The Getinge Assured Helix Test Formaldehyde is a reusable device accompanied with single use chemical indicator strips. The indicator strips will change color from red to green when exposed to appropriate sterilization conditions. The adhesive indicator strips can be used for record keeping.

Physical Characteristics

- Dimensions strip: L 77mm x W 6mm
- Packaging: 1 Helix device and 250 indicators

Environmental Limitations

Storage: Store at normal room temperature 10°–38°C and 10-60% Relative Humidity away from oxidizing agents. Protect from moisture and excess humidity.

Shelf Life

24 months. The expiry date is printed on the product packaging.

Ordering Information

Getinge Assured

Helix Test Formaldehyde

Order Number

504053100